

		CAL	IBRATIO	V CERTI	FICATE				
CERT	IFICATE NO:	SBS/CL/23/13145		MI	ECHANICAL	Pa	ge No:1 of 1		
Issue	Date	· ·	06-10-	2023		and the second s	The second section is a second section of the second		
SRF No & Date			SRF/2	SRF/23/00682-0002 & 05-10-2023					
Receipt Date			05-10-2023						
Calibration Date			05-10-2023						
Calibration Due			04-10-	04-10-2024					
Custo	mer Name & Ad	dress							
SOVE	RNMENT PRIMA	ARY HEALTH CENTRE,							
PERI	APATTINAM-62	3523,RAMANATHAPUR	AM DISTRICT.						
		C	etails of Device U	Inder Calibrat	ion (DUC)		and the second		
Descr	iption :	CENTRIFUGE	Make	& Model		MICROTEK & NA			
Range : 3500 F		3500 RPM	Sr. No		:	LB002255			
Resolution : 1		1 RPM	Identification No			NA			
DUC (Condition :	Satisfactory	Locati	on	- :	LABORATORY			
		Environmenta	l Conditions & St	andard Opera	ting Procedu	re Details			
Environmental Details Temperatu		ure: 25.4°C	Relative Hui	midity	idity 54% Rh				
Calibration Procedure No SBS/CP/M			/IL/04	J04 Calibration		one at ONSITE			
			Reference S	tandards Det	alls		1.		
S.No	o Description		Make/ SI No:	Make/ SI No:		Certificate No			
Digital Tachometer		LINE SEIKI / 1	LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013				


CALIBRATION RESULTS

Ş.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	1000	998.6	1.4	4.2	
2	2000	1998.6	1.4	4.2	
3	3000	2998.7	1.3	4.2	

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By (Calibration Engineer) M.DINESH





SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.