



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. UPHC, Mayiladuthurai

Constituent Group Chemistry III

PT Item Lyophilized human serum based

Lab No

Date of Result Entered :

Date of Report Published :

11310

20/03/2023

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation )	74	92.52	7.71	7.13	88 mg/dL	-0.63	1.66
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	70	1.32	7.52	0.10	1.3 mg/dL	-0.20	0.02
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	148	111.55	4.29	4.78	112 mg/dL	0.09	0.79
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within $\pm 1.01$ to $\pm 2.00$				Good.						
Within $\pm 2.01$ to $\pm 2.99$				Accept with caution. Warning Signal.						
Beyond $\pm 3.0$				Unacceptable performance. Action Signal.						

## LAB ADDRESS :

Govt. UPHC, Mayiladuthurai  
THE MEDICAL OFFICER, MAYILADUTHURAI  
Mayiladuthurai  
TAMILNADU609001

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Medical Officer,  
Urban Primary Health Centre,  
Mayiladuthurai

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore