



Trust & Care



CALIBRATION REPORT	
Report No: HBMS/CAL/22-23	Calibration date: 22.09.2023
	Calibration due: 21.03.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, K.R.THOPPUR,SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	MICROLUX
II Model	100µl - 1000µl
III Serial No	PHC/KT/PP/002

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel I (Volume)		
	100µl	500µl	1000µl
1	100.05	500.08	999.97
2	100.06	500.07	1000.02
3	100.01	500.05	999.96
4	99.97	499.99	1000.02
5	100.01	500.08	1000.09
6	100.02	499.95	1000.04

For Heart Beat Medical Systems,

(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016  
 GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
 88833 15662

No.42/1, Navaindia Road, Peelamedu,  
 CBE - 04, hbmedicalsystem@gmail.com