



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JANUARY 2023



Lab Name Govt. Block PHC, Kandamanadi Lab No 7277  
 Constituent Group Chemistry III Date of Result Entered : 20/01/2023  
 PT Item Lyophilized human serum based Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1890	119.87	6.38	7.65	121 mg/dL	0.15	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1865	2.21	10.02	0.22	2.6 mg/dL	1.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1794	112.14	6.03	6.76	113 mg/dL	0.13	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kandamanadi  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.  
 Kandamanadi  
 TAMILNADU605401

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
 மருத்துவ ஆய்வகம்,  
 சி.பி. ஆரம்ப சுகாதார நிலையம்  
 கண்டமனாடி, விருதுநகர் மாவட்டம்



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - FEBRUARY 2023



Lab Name Govt. Block PHC, Kandamanadi

Lab No 7277

Constituent Group Chemistry III

Date of Result Entered : 21/02/2023

PT Item Lyophilized human serum based

Date of Report Published : 08/03/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	155.11	6.56	10.18	159 mg/dL	0.38	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	3.40	9.00	0.31	3.7 mg/dL	0.98	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1788	112.79	5.23	5.90	120 mg/dL	1.22	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kandamanadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.  
Kandamanadi  
TAMILNADU605401

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Xp*  
மருத்துவ அலுவலர்,  
அரசு ஆரம்ப சுகாதார நிலையம்  
கண்டமானடி, விழுப்புரம் மாவட்டம்



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. Block PHC, Kandamanadi  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 7277  
Date of Result Entered : 19/03/2023  
Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	81 mg/dL	-1.54	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	115 mg/dL	0.70	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Kandamanadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.  
Kandamanadi  
TAMILNADU605401

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
மருத்துவ அலுவலர்,  
அரசு ஆரம்ப சுகாதார நிலையம்  
கண்டமானடி, விழுப்புரம் மாவட்டம்



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name Govt. Block PHC, Kandamanadi  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 7277  
Date of Result Entered : 20/04/2023  
Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.11 mg/dL	-0.29	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	116 mg/dL	0.76	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Kandamanadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.  
Kandamanadi  
TAMILNADU605401

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
மருத்துவ அலுவலர்,  
அரசு ஆரம்ப சுகாதார நிலையம்  
கண்டமானடி, விழுப்புரம் மாவட்டம்



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. Block PHC, Kandamanadi Lab No 7277  
Constituent Group Chemistry III Date of Result Entered : 20/05/2023  
PT Item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	189 mg/dL	0.86	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	120 mg/dL	0.93	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kandamanadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.  
Kandamanadi  
TAMILNADU605401

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Hx*  
மருத்துவ அலுவலர்,  
அரசு ஆரம்ப சுகாதார நிலையம்  
கண்டமானடி, விழுப்புரம் மாவட்டம்



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. Block PHC, Kandamanadi Lab No 7277  
Constituent Group Chemistry III Date of Result Entered : 20/06/2023  
PT item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	309 mg/dL	1.51	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.9 mg/dL	1.93	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	107 mg/dL	-1.33	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Kandamanadi  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.  
Kandamanadi  
TAMILNADU605401

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
மருத்துவ கல்வியல்,  
அரசு ஆரம்ப சுகாதார நிலையம்  
கண்டமாண்டி, விழுப்புரம் மாவட்டம்