



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 03.10.2023
	Calibration Valid: 02.04.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, URBAN PRIMARY HEALTH CENTRE, PALAYAPETTAI, EDAPPADI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	MICROLUX
II Model	5µl - 50µl
III Serial No	UPHC/PYI/PIP/001
IV Location	LABORATORY

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel I (Volume)		
	05µl	20µl	50µl
1	05.05	20.08	50.97
2	05.06	20.07	20.02
3	05.01	20.05	20.96
4	05.97	20.99	20.02
5	05.01	20.08	20.09
6	05.02	20.95	20.04

For Heart Beat Medical Systems,

  
(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016  
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
88833 15662

No.42/1, Navaindia Road, Peelamedu,  
CBE - 04, hbmedicalsystem@gmail.com