



CALIBRATION CERTIFICATE

Page No 1 of 1

Certificate No: SBS/CL/21/1390

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
GOVINDHAMPALAYAM-639008, KARUR DISTRICT.

SRF No.	SRF/23/00737-0004
SRF Date	07-10-2023
Date of Receipt	06-10-2023
Date of Calibration	07-10-2023
Due Date for Calibration	06-10-2024
Issue Date	07-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	MICROLUX
Range	10-100µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.6	Humidity (%RH)	54	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CI/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (s)	
1	10	9.96	9.95	9.98	9.96	9.97	9.97	-0.03	0.01	0.47	
		9.98	9.96	9.97	9.98	9.95					
2	50	49.86	49.85	49.87	49.86	49.85	49.91	-0.09	0.05	0.47	
		49.95	49.96	49.97	49.95	49.95					
3	100	99.96	99.97	99.98	99.96	99.98	99.96	-0.04	0.01	0.47	
		99.95	99.96	99.94	99.95	99.96					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

M. Madhavan
(Calibration Engineer)
MADHAVAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13389 MEDICAL DEVICES Page No:1 of 1

Issue Date	07-10-2023
SRF No & Date	SRF/23/00737-0003 & 07-10-2023
Receipt Date	07-10-2023
Calibration Date	07-10-2023
Calibration Due	06-10-2024

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE,
 GOVINDHAMPALAYAM-639008,KARUR DISTRICT.

Details of Device Under Calibration (DUC)

Description	: ELECTRICAL SAFETY (MICROSCOPE)	Make & Model	: OLAMPUS & NA
Range	: MULTI	Sr. No	: 111/076
Resolution	: MULTI	Identification No	: NA
DUC Condition	: SATISFACTORY	Location	: LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024


ELECTRICAL SAFETY

RESULTS


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 99	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 197	Uncertainty in % (±) 9.2
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 218	Uncertainty in % (±) 8.8

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 K.SATHYAMOORTHY



Authorised Signatory

 Technical Manager
 C.SHANMUGARAJ

Chief Executive