

CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	
Calibration Date: 30.09.2023	
Calibration Valid: 29.03.2024	
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT.PRIMARY HEALTH CENTRE, THIRMMARAVUTHAMPATTY,NAMAKKAL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I	Manufactured by MICROLUX
II	Model -----
III	Value 100µl
IV	Serial No GPHC/TMTY/PIP/003

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	100µl
1	100.05
2	100.06
3	100.01
4	99.97
5	100.01
6	100.02

For Heart Beat Medical Systems,

(Authorized signatory)