



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name: Govt. UPHC, Raja Street      Lab No: 10908  
 Constituent Group: Chemistry III      Date of Result Entered: 20/03/2023  
 PT Item: Lyophilized human serum based      Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	102 mg/dL	1.54	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1701	110.93	5.23	5.80	115 mg/dL	0.70	0.28
SDI Range      Interpretation										
				Within -1.00 to +1.00      Excellent.						
				Within ±1.01 to ±2.00      Good.						
				Within ±2.01 to ±2.99      Accept with caution. Warning Signal.						
				Beyond ±3.0      Unacceptable performance. Action Signal.						

LAB ADDRESS :  
 Govt. UPHC, Raja Street  
 THE MEDICAL OFFICER, RAJA STREET, NEAR CORPORATION CAR PARKING TOWN HALL  
 Coimbatore  
 TAMILNADU641001

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential