

LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Sengal
 Lab No: 12297
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

| No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI |
|----|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|
| | | | | | | CV | SD | | |
| | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 145 mg/dL | -3.4 |
| | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.63 mg/dL | -1.09 |
| | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 116 mg/dL | 0.47 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Outside ±3.0 | Unacceptable performance. Action Signal. |

| | | | |
|--|------------------------------------|---------------------------------|--|
| View Lab Monthly Summary | Submit Lab Results | Modify Methods | View Lab Monthly Summary |
| Sign-out | Home | Change Password | |