



**Vanavil Calibrations**

Make It Simple, But Significant !

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE , KARUPPUR - 636 012.		<b>Certificate No:</b>	VCPL/CC/23-24/39-002
		<b>Date of Calibration:</b>	05-10-2023
		<b>Date of issue:</b>	06-10-2023
		<b>Date of Receipt:</b>	05-10-2023
		<b>Next cal Date:</b>	04-10-2024
<b>Details of Device Under Calibration</b>			
<b>Description:</b>	MICROPIPETTE	<b>Make:</b>	THERMO SCIENTIFIC
<b>Range:</b>	10 - 100 µl	<b>Model/Type:</b>	FINNPIPETTE F3
<b>Resolution:</b>	0.2 µl	<b>Condition of DUC</b>	GOOD
<b>Serial Number</b>	RW090137	<b>Location:</b>	LABORATORY
<b>ID number</b>	KGPHC/MP/01	<b>Cal. At:</b>	LAB
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
<b>Temperature</b>	23.5 °C	<b>Calibration Method Used</b>	ISO 8655:2002 PART-6
<b>Humidity</b>	51 % Rh	<b>Cal procedure no:</b>	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024


### RESULT OF CALIBRATION


S.NO	Duc Readings ( µl )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	10.0	10.02	-0.02
2	50.0	49.96	0.04
3	100.0	99.91	0.09

### Remarks:


- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
**R.GOPALAKRISHNAN**  
(CALIBRATION ENGINEER)



Authorised by:

  
**T. SUDHAKAR,**  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

**Vanavil Calibrations Pvt. Ltd**

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishj Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 6

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

vanavilcalibrations@gmail.com @ www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com