

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Kandamanadi				Details About Robust Analysis					
Lab No	7277				Detail About Monthly Summary					
Month	July				Detail about SDI					
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data					All Analyser Result		Print			
					Date of Result Entered : 20/07/2023					
					Date of Report Published : 07/08/2023					
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	98 mg/dL	-1.87	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.5 mg/dL	1.91	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 மருத்துவ அலுவலர்,
 அரசு ஆரம்ப சுகாதார நிலையம்
 கண்டமானடி, விழுப்புரம் மாவட்டம்