



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. Block PHC, Manickampalayam Lab No 4604
 Constituent Group Chemistry III Date of Result Entered : 20/09/2023
 PT item Lyophilized human serum based Date of Report Published : 05/10/2023

Sl. No.	Analyte	Method / Principle / Reagent	Analyte Name	No of Participants	DV	Participants		Your Value	SDI	Tf
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	318.9 mg/dL	0.54	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.33 mg/dL	0.38	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	111.2 mg/dL	-0.56	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Manickampalayam

MEDICAL OFFICER, PULLAGOUNDAMPATTY POST, THIRUCHENGODE TALUK, NAMAKKAL DISTRICT

Manickampalayam

TAMILNADU637202

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

R. Karan...
 MEDICAL OFFICER,
 GOVT. PRIMARY HEALTH CENTRE
 MANICKAMPALAYAM POST 637 202
 Thiruchengode Taluk, Namakkal Dist.