



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name
 Constituent Group
 PT item

Govt. Block PHC, Manickampalayam
 Chemistry III
 Lyophilized human serum based

Lab No

4604

Date of Result Entered :

21/04/2023

Date of Report Published :

06/05/2023

Sl. No.	Parameter	Method	Manufacturer	Observed	Target	Delta	Delta 2	Delta 3	Delta 4	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124.01 mg/dL	0.63	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.04 mg/dL	-0.69	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109.2 mg/dL	-0.42	0.28

SDI Range

Excellent.

Good.

Accept with caution. Warning Signal.

Unacceptable performance. Action Signal.

Within -1.00 to +1.00

Within ± 1.01 to ± 2.00

Within ± 2.01 to ± 2.99

CHENGODE TALUK, NAMAKKAL DISTRICT



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name Govt. Block PHC, Manickampalayam Lab No 4604
Constituent Group Chemistry III Date of Result Entered : 21/04/2023
PT item Lyophilized human serum based Date of Report Published : 06/05/2023

Sl. No.	Test Name	Method	Instrument	Result	Mean	SD	CV	Reference Range	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124.01 mg/dL	0.63	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.04 mg/dL	-0.69	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109.2 mg/dL	-0.42	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Manickampalayam
MEDICAL OFFICER, PULLAGOUNDAMPATTY POST, THIRUCHENGODE TALUK, NAMAKKAL DISTRICT
Manickampalayam
TAMILNADU637202

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Blm
MEDICAL OFFICER,
GOVT. PRIMARY HEALTH CENTRE
MANICKAMPALAYAM-537 202
Thiruchengode Tk. Namakkal Dt



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



4604

Lab Name Govt. Block PHC, Manickampalayam
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No

Date of Result Entered :

Date of Report Published :

21/04/2023

06/05/2023

Sl. No.	Parameter	Method	Instrument	Result	Target	Delta	Delta-2	Delta-3	Delta-4	Delta-5
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124.01 mg/dL	0.63	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.04 mg/dL	-0.69	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109.2 mg/dL	-0.42	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Manickampalayam
MEDICAL OFFICER, PULLAGOUNDAMPATTY POST, THIRUCHENGODE TALUK, NAMAKKAL DISTRICT
Manickampalayam
TAMILNADU637202

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Blent
MEDICAL OFFICER,
GOVT. PRIMARY HEALTH CENTRE
MANICKAMPALAYAM-537 202
Thiruchengode Tk. Namakkal Dt