



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name	Govt. Block PHC, Manickampalayam	Lab No	4604
Constituent Group	Chemistry III	Date of Result Entered	20/06/2023
PT item	Lyophilized human serum based	Date of Report Published	04/07/2023

Sl. No.	Parameter	Method	Instrument	1	2	3	4	5	6	7	8
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	290 mg/dL	0.23	0.69	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116.2 mg/dL	0.49	0.24	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Manickampalayam

MEDICAL OFFICER, PULLAGOUNDAMPATTY POST, THIRUCHENGODE TALUK, NAMAKKAL DISTRICT

Manickampalayam

TAMILNADU637202

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 End of Report

S. Babu
MEDICAL OFFICER,
GOVT. PRIMARY HEALTH CENTRE,
MANICKAMPALAYAM-537 202,
Tiruchengode Taluk, Namakkal Dt