

Sep-2023


L J CHART- Creatinine

Mean : 1.55mgs/dl

CMC-IQC

Govt. Primary Health Centre - Keeripatti - phc

Range : 1.30 - 1.80mgs/dl



Max Value	1.800
Mean Value	1.550
2SD (1SD x 2)	0.250
2SD for Analyte (Mean+2SD)	1.800

Max Value	1.800
Mean Value	1.550
1SD (Mean-Max)/2	0.125
1SD for Analyte (Mean+1SD)	1.675

Mean
1.55

Min Value	1.300
Mean Value	1.550
(-1)SD (Mean-Min)/2	0.125
(-1)SD for Analyte (Mean-2SD)	1.425

Min Value	1.300
Mean Value	1.550
(-2)SD (1SD x 2)	0.250
(-2)SD for Analyte (Mean-2SD)	1.300

Result	1.66	1.675	1.60	1.55	1.50	1.475	1.55	1.60	1.675	1.60	1.50	1.475	1.55	1.60	1.50					
Date	4	5	6	7	12	13	14	19	20	21	26	27	28	29	30					

Precicon 10 *Truh*

MEDICAL OFFICER
Govt. Primary Health Centre.
KEERIPATTY (Po).
Attur (Tk), Salem (Dt) 636 107