

CHRISTIAN MEDICAL COLLEGEDEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2021**CHRISTIAN MEDICAL COLLEGE**DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2021**CHRISTIAN MEDICAL COLLEGE**DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2021

Lab Name: Govt. Block PHC, Marikampally

Complaint Group: Chemistry-B

PT Item: Lyophilized human serum based

Test	Ref. Range	Result (1 Sample)	Automated / Fully Automated / Partially Automated / Manual	Unit	Interpretation
1. GLUCOSE-B	100-140 mg/dL	123.45	1.04	7.21 mg/dL	Acceptable
2. CREATININE-B	60-120 mg/dL	87.68	0.16	2.08 mg/dL	Acceptable
3. CHOLESTROL-B (CHOD-PAP-B)	170-220 mg/dL	191.81	0.19	2.78 mg/dL	Acceptable

SDS Range

Range	Interpretation
Within 100 to 1100	Excellent
Within 111 to 119	Good
Within 120 to 139	Accept with caution. Warning Signal
Beyond 140	Unacceptable performance. Action Signal

Lab Address:

Govt. Block PHC, Marikampally
MEDICAL OFFICER: FULLAGOLUNDAMPARTY POST, THIRUCHENGODE TALUK, NAMAKKAL DISTRICT
Marikampally
TAMIL NADU 641702Coordinator Contact Details:
Email-Id: jgmsc@outlook.com.in
Contact Number: 9444223102Homogeneity and Stability of the sample is passed.
Data in CMC EQAS report is confidential.
CMC EQAS does not contact any companies
-----End of Report-----**MEDICAL OFFICER,**
GOVT. PRIMARY HEALTH CENTRE,
MANIKKAPALAYAM 641262,
THIRUCHENGODE TALUK, NAMAKKAL**CHRISTIAN MEDICAL COLLEGE**
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2021**CHRISTIAN MEDICAL COLLEGE**
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2021**CHRISTIAN MEDICAL COLLEGE**
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2021

Lab Name: Govt. Block PHC, Marikampally

Complaint Group: Chemistry-B

PT Item: Lyophilized human serum based

Test	Ref. Range	Result (1 Sample)	Automated / Fully Automated / Partially Automated / Manual	Unit	Interpretation
1. GLUCOSE-B	100-140 mg/dL	129.22	0.33	8.21 mg/dL	Acceptable
2. CREATININE-B	60-120 mg/dL	108.60	0.07	2.08 mg/dL	Acceptable
3. CHOLESTROL-B (CHOD-PAP-B)	170-220 mg/dL	191.86	0.12	2.48 mg/dL	Acceptable

SDS Range

Range	Interpretation
Within 100 to 1100	Excellent
Within 111 to 119	Good
Within 120 to 139	Accept with caution. Warning Signal
Beyond 140	Unacceptable performance. Action Signal

Coordinator Contact Details:
Email-Id: jgmsc@outlook.com.in
Contact Number: 9444223102Homogeneity and Stability of the sample is passed.
Data in CMC EQAS report is confidential.
CMC EQAS does not contact any companies
-----End of Report-----**MEDICAL OFFICER,**
GOVT. PRIMARY HEALTH CENTRE,
MANIKKAPALAYAM 641262,
THIRUCHENGODE TALUK, NAMAKKAL**CHRISTIAN MEDICAL COLLEGE**
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - OCTOBER 2021**CHRISTIAN MEDICAL COLLEGE**
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2021

Lab Name: Govt. Block PHC, Marikampally

Complaint Group: Chemistry-B

PT Item: Lyophilized human serum based

Test	Ref. Range	Result (1 Sample)	Automated / Fully Automated / Partially Automated / Manual	Unit	Interpretation
1. GLUCOSE-B	100-140 mg/dL	121.45	0.34	7.27 mg/dL	Acceptable
2. CREATININE-B	60-120 mg/dL	87.69	0.15	2.08 mg/dL	Acceptable
3. CHOLESTROL-B (CHOD-PAP-B)	170-220 mg/dL	191.80	0.17	2.44 mg/dL	Acceptable

SDS Range

Range	Interpretation
Within 100 to 1100	Excellent
Within 111 to 119	Good
Within 120 to 139	Accept with caution. Warning Signal
Beyond 140	Unacceptable performance. Action Signal

Coordinator Contact Details:
Email-Id: jgmsc@outlook.com.in
Contact Number: 9444223102Homogeneity and Stability of the sample is passed.
Data in CMC EQAS report is confidential.
CMC EQAS does not contact any companies
-----End of Report-----