

				CA	LIBRA	ATIO	N CEF	RTIF	-IC	<u>ATE</u>	
CERTIFICATE NO: SBS/CL/23/13389					MEDICAL DEVICES					Page No:1 of 1	
Issue Date					07-10	0-2023					
SRF No & Date					SRF/23/00737-0003 & 07-10-2023						
Receipt Date					07-10-2023						
Calibration Date					07-10	07-10-2023					
Calibration Due					06-10	06-10-2024					
Cust	omer Nam	e & Ad	dress								
GOV	ERNMENT	PRIMA	ARY HE	ALTH CENTE	RE,						
GOV	INDHAMPA	ALAYA	M-63900	B,KARUR DI	STRICT.						
					Details of	of Device	Under Calib	ration	(DUC	()	
	Merchanic Committee		1900 1000 1000	TRICAL SAFE	TY						
	ription	:	(MICR	OSCOPE)		Make	& Model		:	OLAMPUS & NA	
Range		:	MULT	ı		Sr. No)		:	I11/076	
Resolution		:	MULT	1		Identi	fication No		:	NA	
DUC	Condition	:	SATIS	FACTORY		Locat			:	LABORATORY	
				Enviro	onmental C	onditions	& Calibrati	ion Pro	cedu	re Details	
Environmental Details Temperature			25.6°C Relative		Relative Hu	lumidity		52% RH			
Calibration Procedure No SBS/CP/MD			29 Calibration		done at ONSITE		ONSITE				
					R	eference	Standards I	Details			
S.No	o Description			Make/ SI No:		Certificate No			Validity		
1 Electrical Safety Analyser			Rigel Medical & 44L-1059		M-230809-16-4			10-08-2024			

		ELECTRICAL SAFETY						
RESULTS								
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)					
1	Insulation Resisitance	Measured values in $M\Omega$	Uncertainty in % (±)					
	>20MΩ	99	13.92					
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)					
	<5000µAfor B,BF,CF	197	9.2					
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)					
	<500µAfor B,BF,CF	218	8.8					

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.