

## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO:</b>	SBS/CL/23/12989	<b>MECHANICAL</b>	<b>Page No:1 of 1</b>
Issue Date	06-10-2023		
SRF No & Date	SRF/23/00616-0005 & 05-10-2023		
Receipt Date	05-10-2023		
Calibration Date	05-10-2023		
Calibration Due	04-10-2024		

**Customer Name & Address**  
 GOVERNMENT PRIMARY HEALTH CENTRE(MAILAM BLOCK),  
 MANNAMPOONDI-604207,VILLUPURAM DISTRICT.

### Details of Device Under Calibration (DUC)

Description : CENTRIFUGE	Make & Model : B.D INSTRUMENTATION & BDI-152
Range : 3500 RPM	Sr. No : NA
Resolution : 1 RPM	Identification No : NA
DUC Condition : Satisfactory	Location : LABORATORY

### Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

### Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & DT-2238	JRPM-CCTR-A&S-2022-0036	03-10-2023

### CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY ( ± )
	<b>RPM</b>	<b>RPM</b>	<b>RPM</b>	<b>%</b>
1	1000	999.9	0.1	4.2
2	2000	1998	2	4.2
3	3000	2999	1	4.2

### REMARKS


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
 (Calibration Engineer)  
 M.DINESH



Authorised Signatory

  
 Quality Manager  
 (C.SIVABALAN)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.