



## CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13388	MECHANICAL	Page No:1 of 1	
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00737-0002 & 07-10-2023			
Receipt Date	07-10-2023			
Calibration Date	07-10-2023			
Calibration Due	06-10-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, GOVINDHAMPALAYAM-639008,KARUR DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	: CENTRIFUGE	Make & Model	: LABOTECH & NA	
Range	: 3500 RPM	Sr. No	: MK230X1	
Resolution	: 1 RPM	Identification No	: NA	
DUC Condition	: Satisfactory	Location	: LABORATORY	
<b>Environmental Conditions &amp; Standard Operating Procedure Details</b>				
Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

### CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY ( ± )
	RPM	RPM	RPM	%
1	1000	999.7	0.3	4.2
2	2000	1999.8	0.2	4.2
3	3000	2999.9	0.1	4.2

### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

*K.S.*  
(Calibration Engineer)  
K.SATHYAMOORTHY



Authorised Signatory

*C. Sivalan*  
 Quality Manager  
(C.SIVABALAN)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.