

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13391 Page. No: 1 of 1

Customer Name & Address		SRF No.	
GOVERNMENT PRIMARY HEALTH CENTRE, GOVINDHAMPALAYAM-639008,KARUR DISTRICT.		SRF/23/00737-0005	
		SRF Date	
		07-10-2023	
		Date of Receipt	
		06-10-2023	
		Date of Calibration	
07-10-2023			
Due Date for Calibration		06-10-2024	
Issue Date		07-10-2023	

Details of Unit Under Calibration			
Description	MICRO PIPETTE	Make	MICROLUX
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY


Environmental Condition				Calibration Method Used	
Temperature (°C)	23.5	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08


Standard Used						
Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.94	99.96	99.94	99.92	99.93	99.95	-0.05	0.02	0.47
		99.94	99.95	99.96	99.97	99.98				
2	500	499.85	499.86	499.87	499.89	499.86	499.87	-0.13	0.02	0.47
		499.86	499.87	499.89	499.85	499.89				
3	1000	999.96	999.97	999.98	999.96	999.95	999.96	-0.04	0.01	0.47
		999.98	999.94	999.96	999.95	999.95				

- Remarks**
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - The user should determine the suitability of the instrument for its intended use.
 - The recalibration interval should be determined on the user requirement.
 - The results stated in this certificate relate only to the item calibrated.
 - Equipment used for Calibration were calibrated & traceable to National & International Standards
 - The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
 - Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By, 
(Calibration Engineer)
M.ADHIBAN



Authorised by:

(Quality Manager/Chief Executive)
C.SIVABALAN