

| CALIBRATION REPORT | |
|---|--|
| Report No : HBMS/CAL/23-24 | Calibration Date : 23.09.2023 |
| | Calibration valid : 22.03.2024 |
| 1.1.CUSTOMER DETAILS: | |
| Name and Address of the organization | M/s, PRIMARY HEALTH CENTRE, PETHANAICKENPALAYAM, SALEM. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | |
| EQUIPMENT NAME | MICRO PIPETTE |
| i Manufactured by | THERMOFISHER |
| ii Model | FINNPIPETTE |
| iii Value | 10 µl -100 µl |
| iv Serial No | RW07903 |

Calibration standards used :

| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Calibration test Report.

| Number of measurement | Channel 1 (Volume) | | |
|-----------------------|--------------------|-------|--------|
| | 10µl | 50µl | 100µl |
| 1 | 10.03 | 50.06 | 100.14 |
| 2 | 10.04 | 50.09 | 100.12 |
| 3 | 10.05 | 50.02 | 100.10 |
| 4 | 10.09 | 50.08 | 100.08 |
| 5 | 10.00 | 49.94 | 100.09 |
| 6 | 10.08 | 50.07 | 100.12 |

For Heart Beat Medical Systems,

(Authorized signatory)