



# RAJA MEDICAL SYSTEMS

No.15, V.O.C STREET, COIMBATORE - 641 011. MOBILE : 96263 13796.

## Calibration Protocol & Report for Medical Equipment

Equipment : **CENTRIFUGE**  
MAKE : REMI  
MODEL : R-8C

PROTECT CLASS /TYPE : Risk Class I  
S. No. : ZFAN-03119  
Location : AUPHC,ERODE

S.No	Measurement Parameters		Status	Remarks
1	<b>Electrical Test</b>	(a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS  With in Ranges Well
2	<b>Physical Condition Test</b>	(a) Cabinet & Visible damages, if any (b) Front Panet/Top lid, buckets, motor asembly & power cord (c) Risk of mechanical damages, if any	No OK No	Checked OK
3	<b>Performance Test</b>	(a) On /Off switch (b) Speed regulator switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK Accurate OK No Well Well	Ok Perfect Perfect
4	<b>Calibration Test</b>	(a) Vibration test (b) Speed regulation gradual hike/wobbling test (c) Running test with load (d) Display checkup (e) Timer test	Done Ok	Perfect
<b>Formulated By:</b> Einstein Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>	
<b>Endorsed By :</b> Research team head		<b>RAJA MEDICAL SYSTEMS</b> No.15, V.O.C. STREET, COIMBATORE - 641 011.		
Reviewed On: 29.10.22 Next Review : 28.10.23				