

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Nelliyalam  
 Lab No 12487  
 Month July  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

[Click on the analyte to view Graphical Data](#)

| Sl.No                           | Analyte        | Method / Principle Name               | Analyser Name                                | No of Participants    | DV     | Participants CV | SD   | Your Value | SDI   | U    |  |  |
|---------------------------------|----------------|---------------------------------------|--|-----------------------|--------|-----------------|------|------------|-------|------|--|--|
| 1                               | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1876                  | 109.35 | 5.55            | 6.07 | 113 mg/dL  | 0.60  | 0.28 |  |  |
| 2                               | UREA III       | Urease UV / GLDH III                  | Robonik ( Semi automatic / Fully Automatic ) | 184                   | 51.41  | 11.14           | 5.73 | 44 mg/dL   | -1.29 | 0.84 |  |  |
| 3                               | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1850                  | 5.83   | 6.00            | 0.35 | 5.9 mg/dL  | 0.20  | 0.02 |  |  |
| 4                               | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1776                  | 113.00 | 4.12            | 4.66 | 109 mg/dL  | -0.86 | 0.22 |  |  |
| <b>SDI Range Interpretation</b> |                |                                       |  |                       |        |                 |      |            |       |      |  |  |
|                                 |                |                                       |  | Within -1.00 to +1.00 |        |                 |      |            |       |      | Excellent.                               |  |
|                                 |                |                                       |  | Within ±1.01 to ±2.00 |        |                 |      |            |       |      | Good.                                    |  |
|                                 |                |                                       |  | Within ±2.01 to ±2.99 |        |                 |      |            |       |      | Accept with caution. Warning Signal.     |  |
|                                 |                |                                       |  | Beyond ±3.0           |        |                 |      |            |       |      | Unacceptable performance. Action Signal. |  |

0  
 100