

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Pattukottai
 Lab No 9357
 Month March
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI


All Analyser Result Print

Date of Result Entered : 19/03/2023

Date of Report Published : 04/04/2023

Sl. No	Parameter	Method	Instrument	Count	Mean	SD	CV	Result	SDI	Interpretation
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.48	8.2	91 mg/dL	-0.07	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	6.82	1.2 mg/dL	-1.07	0.01
3	CHOLESTEROL	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	4.72	114 mg/dL	0.53	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 (பட்டினம்)

Links