

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kolapalli	Details About Robust Analysis
Lab No	12464	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Analyte Name	Method	Analyzer Model	No of Participants	CV	Participants CV	SD	Mean Value	CVI	CVI
1 GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	122 mg/dL	0.35	0.34
2 UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	41 mg/dL	-0.29	0.70
3 CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
4 CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	116 mg/dL	0.76	0.28
SDI Range		Interpretation							
Within -1.00 to +1.00		Excellent.							
Within ±1.01 to ±2.00		Good.							
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.							
Beyond ±3.0		Unacceptable performance. Action Signal.							

மேன்மைக்கான சேவை
சென்னை
தமிழ்நாடு