

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kolapalli
 Lab No 12464
 Month May
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)


[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl. No.	Parameter	Reference Range	Analyzer Name	No. of Participants	DV	CV	SD	Mean	SD	CV	SD	CV	SD
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	178 mg/dL	-0.31	0.44			
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	24 mg/dL	-1.05	0.59			
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01			
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	110 mg/dL	-0.79	0.27			
SDI Range		Interpretation											
Within -1.00 to +1.00		Excellent.											
Within ±1.01 to ±2.00		Good.											
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.											
Beyond ±3.0		Unacceptable performance. Action Signal.											


 மருத்துவ அலுவலர்
 அரசு சுகாதார அலுவலகம்
 கல்கத்தா - 700 001