

**VIEW LAB MONTHLY SUMMARY**

**Lab Name** Govt. UPHC, Pattukottai  
**Lab No** 9387  
**Month** April  
**Year** 2023  
**Constituent Group** Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	117.25 mg/dL	-0.
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	0.09	0.18	2.1 mg/dL	-0
3	CHOLESTROL II	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	111 mg/dL	-4
<b>SDI Range</b>		<b>Interpretation</b>							
Within -1.00 to +1.00		Excellent.							
Within ±1.01 to ±2.00		Good.							
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.							
Beyond ±3.0		Unacceptable performance. Action Signal.							

**Links**

- [Monthly Summary](#)
- [Quality Dashboard](#)
- [Laboratory Overview](#)
- [Quality Summary](#)
- [Change Management](#)
- [News](#)
- [Reports](#)



*for result*  
 மருத்துவ அலுவலர்  
 மருக ஆரம்ப காலதரம்  
 (நகர்ப்புறம்)  
 மருகக் கார்டை ந