

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Pattukotta
 Lab No: 8287
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

Details About Report Analysis
 Details About Monthly Summary
 Details about SDI

All Analyser Result Print

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl. No.	Parameter	Method	Instrument	Reading	Unit	SDI	SDI	SDI	SDI	SDI	SDI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	289.6	mg/dL	0.20	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.2	mg/dL	-2.16	0.01
3	CHOLESTEROL	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1854	112.73	4.40	5.07	118.9	mg/dL	1.02	0.24

SDI Range	Interpretation
Within -1.00 to -1.00	Excellent.
Within =1.01 to =2.00	Good.
Within =2.01 to =2.99	Accept with caution. Warning Signal.
Beyond =3.0	Unacceptable performance. Action Signal.

Links

(Signature)
 உறுதிப்படுத்தப்பட்டது
 அரசு ஆய்வு கक्षाது துறை
 (தரப்புகள்)
 மதுரை