

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Pattukottal
 Lab No: 0387
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the icons to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl. No.	Substance	Method / Principle / Reagent	Analyzer / Method	Result / Value	SD	Particulate / Ca	TOC	Total Hardness / mg/dL	TH	SDI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	117 mg/dL	1.28	0.26
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTEROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.68	121 mg/dL	1.72	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதாய நகரம்
 (நகர்ப்புறம்)
 மருத்துவக்கல்வித் துறை