

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kolapalli
 Lab No 12464
 Month June
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Analyte	Reference Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Mean Value	SD	CV	
1	GLUCOSE III	GOD-POD III	1823	286.63	5.17	14.82	252 mg/dL	-2.34	0.69	
2	UREA III	Urease UV / GLDH III	192	32.69	12.96	4.24	21 mg/dL	-2.76	0.61	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1848	2.57	6.64	0.17	2.1 mg/dL	-2.75	0.01	
4	CHOLESTROL III	CHOD-PAP III	1804	113.73	4.46	5.07	108 mg/dL	-1.52	0.24	
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

DATE: 04/07/2023
 TIME: 11:52 AM
 USER: GNA