



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. PHC, Thevoor

Lab No 12615

Constituent Group Chemistry III

Date of Result Entered : 20/09/2023

PT item Lyophilized human serum based

Date of Report Published : 05/10/2023

| Sl No | Anal           | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |       | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD    |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1870               | 310.86 | 4.82         | 14.98 | 278 mg/dL  | -2.19 | 0.69 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1852               | 1.29   | 8.14         | 0.10  | 1.3 mg/dL  | 0.10  | 0.00 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1783               | 113.62 | 3.80         | 4.32  | 113 mg/dL  | -0.14 | 0.20 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

**LAB ADDRESS :**

Govt. PHC, Thevoor  
THEVOOR, SANKAGIRI  
Thevoor  
TAMILNADU 637104

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*