VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Pattikottai
Lab No	9387
Month	August
Year	2023
Constituent Group	Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result | Print

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	190.88 mg/dL	1.70	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.69 mg/dL	-0.59	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	117.64 mg/dL	0.83	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Scanned with OKEN Scanner