

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. UPHC, Panjupettai**
 Lab No **10805**
 Month **May**
 Year **2023**
 Constituent Group **Chemistry III**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	CV	SD	CV	SD	CV	SD
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	190 mg/dL	0.84	2.53	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.7 mg/dL	-0.41	0.03	
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	105 mg/dL	-1.69	0.94	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Monthly Summary	Detailed Lab Reports	View Lab Monthly Summary
Change Password	Logout	Sign-out

Amulya