

**CALIBRATION CERTIFICATE**

Certificate No: SBS/CL/23/13377 Page No: 1 of 1

<b>Customer Name &amp; Address</b>		<b>SRF No.</b>	SRF/23/00734-0004
GOVERNMENT PRIMARY HEALTH CENTRE, KONDANGIPATTI-639005,KARUR DISTRICT.		<b>SRF Date</b>	07-10-2023
		<b>Date of Receipt</b>	06-10-2023
		<b>Date of Calibration</b>	07-10-2023
		<b>Due Date for Calibration</b>	06-10-2024
		<b>Issue Date</b>	07-10-2023

<b>Details of Unit Under Calibration</b>			
<b>Description</b>	Micro Pipette	<b>Make</b>	THERMO SCIENTIFIC
<b>Range</b>	100-1000µl	<b>Model</b>	FINNPIPETTE F3
<b>Resolution</b>	1µl	<b>Material</b>	PVC
<b>Serial Number</b>	NA	<b>Operating Range</b>	100-1000µl
<b>ID Number</b>	NA	<b>Condition of UUC</b>	Good
<b>Cal. At</b>	Mechanical Lab	<b>Instrument Location</b>	LABORATORY


<b>Environmental Condition</b>				<b>Calibration Method Used</b>	
<b>Temperature (°C)</b>	23.8	<b>Humidity (%RH)</b>	55	<b>National / International Standard</b>	ISO 8655-6:2002
<b>Atmospheric Pressure (mbar)</b>	1006	<b>Water Temperature (°C)</b>	21.6	<b>Cal Procedure No</b>	SBS/CP/ML/08


<b>Standard Used</b>						
<b>SI. No.</b>	<b>Description</b>	<b>ID.No. / SI. No.</b>	<b>Certificate No.</b>	<b>Make/Model</b>	<b>Traceability</b>	<b>Valid till</b>
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


<b>Z Factor: 1.00319</b>											
<b>Result of Calibration in µl</b>											
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.85	99.87	99.85	99.84	99.95	99.89	-0.11	0.04	0.47	
		99.89	99.95	99.90	99.87	99.90					
2	500	499.90	499.95	499.90	499.90	499.95	499.90	-0.10	0.03	0.47	
		499.89	499.88	499.85	499.85	499.90					
3	1000	999.89	999.88	999.89	999.87	999.90	999.91	-0.09	0.03	0.47	
		999.93	999.90	999.91	999.94	999.95					

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By:   
(Calibration Engineer)  
M.RAGUL

  
 Chennai  
 600 032

Authorised by:   
(Quality Manager/Chief Executive)  
C.SIVABALAN