

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Pattukottai  
 Lab No 8387  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

Details About Robonik Analysis

Detail About Monthly Summary

Detail about SDI

Click on the link (if available) to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

| Sl. No. | Parameter Name | Method                                | Unit  | Value | SDI    | Range | Interpretation |
|---------|----------------|---------------------------------------|-------|-------|--------|-------|----------------|
| 1       | GLUCOSE III    | GOD-POD III                           | mg/dL | 1834  | 180.85 | 5.19  | 9.39           |
| 2       | CREATININE III | Jaffes Kinetic - Alkaline Pictate III | mg/dL | 1854  | 1.76   | 7.48  | 0.13           |
| 3       | CHOLESTROL III | CHOD-PAP III                          | mg/dL | 1798  | 114.60 | 5.08  | 5.83           |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

Links

மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 (நகர்ப்புறம்)  
 மட்டுக்காட்டை தஞ்சை