


VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Asanur	Details About Robust Analysis								
Lab No	13138	Detail About Monthly Summary								
Month	May	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data		All Analyser Result	Print							
Date of Result Entered : 19/05/2023										
Date of Report Published : 05/06/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	184 mg/dL	0.33	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.5 mg/dL	-1.97	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	111 mg/dL	-0.62	0.27
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்,
 ஆசனூர்,
 உளுந்துார்பேட்டை வட்டி