

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Panjupetta
 Lab No: 10805
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	Dv.	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95	100 mg/dL	-1.63	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27	6 mg/dL	0.41	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	108 mg/dL	-0.98	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal!
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify details	Modify Analyser	Submit Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Sign out

A ready