

**VIEW LAB MONTHLY SUMMARY**

Lab Name: Govt. UPHC, Panjupettai  
 Lab No: 10805  
 Month: August  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 17/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants			SDI	U
						CV	SD	Yrly Value		
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51	190 mg/dL	1.65	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13	119 mg/dL	901.85	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	1.5 mg/dL	-22.63	0.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Monthly Summary</a>	<a href="#">Modify Profile</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Monthly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Signout</a>

*Handwritten signature*