

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Ramanathapuram
 Lab No 10870
 Month June
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

| Sl No | Analyte | Method Principle Name | Analyzer Name | No of Participants | DV | Participants CV | SD | Your Value | SDI | U |
|-------|----------------|---------------------------------------|---|--------------------|--------|-----------------|-------|------------|-------|------|
| 1 | GLUCOSE III | GOD-POD III | Any Analyser (Automation / Semi Automation) | 65 | 287.55 | 4.50 | 12.95 | 357 mg/dL | 5.36 | 3.21 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Any Analyser (Automation / Semi Automation) | 67 | 2.58 | 6.39 | 0.16 | 2.62 mg/dL | 0.24 | 0.04 |
| 3 | CHOLESTROL III | CHOD-PAP III | Any Analyser (Automation / Semi Automation) | 149 | 114.61 | 4.05 | 4.64 | 112 mg/dL | -0.56 | 0.76 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Home Address | Lab Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |

[Signature]
 04/07/23
MEDICAL OFFICER
 RAMANATHAPURAM
 UPHC
 Government Corporation