



# G MULTI SERVICES

78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.  
 gmultiservicesgobi@gmail.com Cell - 96597 20654  
 Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE., BE (ICE)  
 Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : SEMI AUTO ANALYSER  
 MAKE : ROBONIK  
 MODEL : PRIESTEST TOUCH

PROTECT CLASS / TYPE : Risk Class I  
 S. No. : ATCD3860321RBK  
 Location : GPHC, KOLAPALLI.

S.No	Measurement Parameters	Status	Remarks
1	<b>Electrical Test</b> (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS  With in Ranges Well
2	<b>Physical Condition Test</b> (a) Cabinet & Visible damages, if any (b) Front Panel/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK  No	Checked OK
3	<b>Performance Test</b> (a) On /Off switch (b) switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK OK No Well Well	Ok Perfect Perfect
4	<b>Calibration Test</b> (a) End Point Test (b) Display checkups (c) Filter selection (d) Record OD (e) Calculation (f) Two Point Test (g) Kinetic Test (h) Result	Done OK OK OK Well Perfect	Perfect
<b>Formulated By:</b> Edition Research Laboratory		<b>Issued By:</b>	
<b>Endorsed By:</b> Research team head		<b>User Department:</b>	
Reviewed On: 13.09.23 Next Review : 12.09.24			