



ORION DIAGNOSTICS

308, Shooting Range West Side. Madurai Road, Poovani.
Srivilliputtur - 626 125. Virudhunagar Dist.
GSTIN : 33CIGPS9450J2ZR
Cell : 93644 33099, 93600 33099

Certificate of Calibration

Date :

Model Name: Prietest Touch

Date of Calibration: 29.12.2022

Sr. No.: AT3010413RBK

Due Date of Calibration: 28.12.2023

The Medical Officer,
Government Primary Health Centre,
Puthiyamputhur.

Calibrated by:

Name: K.SARAVANAN.

Sign: 

Date: 29.12.22

Approved by:

Name: P.SRINIVASAN.

Sign: 

Date: 29.12.22



ORION DIAGNOSTICS

308, Shooting Range West Side. Madurai Road, Poovani.

Srivilliputtur - 626 125. Virudhunagar Dist.

GSTIN : 33CIGPS9450J2ZR

Cell : 93644 33099, 93600 33099

Model Name: Prietest Touch

Sr.No: AT3010413RBK

Wavelength	Film type	Range	Absorbance Obtained
340 nm	HP	0.333-0.361	0.334
	DP	1.164-1.253	1.167
	TP	1.658-1.782	1.669

Date

	Film type	Range	Absorbance Obtained
405 nm	DP	0.661-0.708	0.661
	3DP	1.943-2.098	1.950
	TP	1.006-1.077	1.015
	2TP	1.932-2.099	1.956

Wavelength	Film type	Range	Absorbance Obtained
510 nm	HP	0.161-0.173	0.165
	DP	0.647-0.695	0.659
	3DP	1.910-2.081	1.911
	TP	0.969-1.036	0.981

Wavelength	Film type	Range	Absorbance Obtained
545nm	HP	0.154-0.165	0.158
	SP	0.304-0.321	0.316
	DP	0.595-0.629	0.599
	TP	0.903-0.956	0.916

Wavelength	Film type	Range	Absorbance Obtained
578nm	DP	0.576-0.616	0.581
	3DP	1.657-1.857	1.671
	TP	0.867-0.922	0.873
	3TP	2.549-2.738	2.556

Wavelength	Film type	Range	Absorbance Obtained
630nm	HP	0.162-0.176	0.163
	DP	0.653-0.712	0.658
	TP	1.043-1.116	1.050
	2TP	2.009-2.166	2.020

Calibrated By: K.SARAVANAN

Signature:

Date: 29.12.22

Approved By: P.SRINIVASAN

Signature:

Date: 29.12.22