


VIEW LAB MONTHLY SUMMARY

| Lab Name | Govt. PHC, Asanur | Details About Robust Analysis | | | | | | | | |
|---|-------------------|---------------------------------------|--|--------------------|--------|-------------------------------------|-----------------|-----------------------|-------|------|
| Lab No | 13138 | Detail About Monthly Summary | | | | | | | | |
| Month | August | Detail about SDI | | | | | | | | |
| Year | 2023 | | | | | | | | | |
| Constituent Group | Chemistry III | | | | | | | | | |
| Click on the analyte to view Graphical Data | | | | | | All Analyser Result | | Print | | |
| Date of Result Entered : 19/08/2023 | | | | | | | | | | |
| Date of Report Published : 01/09/2023 | | | | | | | | | | |
| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants CV | Participants SD | Your Value | SDI | U |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 176 mg/dL | 0.05 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.7 mg/dL | -0.50 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 106 mg/dL | -1.68 | 0.22 |
| SDI Range | | | Interpretation | | | | | | | |
| Within -1.00 to +1.00 | | | Excellent. | | | | | | | |
| Within ±1.01 to ±2.00 | | | Good. | | | | | | | |
| Within ±2.01 to ±2.99 | | | Accept with caution. Warning Signal. | | | | | | | |
| Beyond ±3.0 | | | Unacceptable performance. Action Signal. | | | | | | | |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |


 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்,
 ஆசனூர்,
 உளுந்துார்பேட்டை வட்டம்.