



### CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 06.10.2023
		Calibration Valid: 05.04.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE, PAKKANADU, SALEM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
<b>i</b>	Manufactured by	REMI
<b>ii</b>	Model	R8C
<b>iii</b>	Serial No	ZELN 42118
<b>iv</b>	Location	LABORATORY

### CALIBRATION WORKSHEET

#### Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1001	1
2	1200	1199	-1
3	1500	1500	0
4	2000	1998	-2

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation:  $\pm 0.5$  seconds

#### Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non – contact ) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

#### Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :



The image shows a handwritten signature in blue ink over a circular purple stamp. The stamp contains the text 'HEARTBEAT MEDICAL SYSTEMS' around the perimeter and 'CBE' in the center.



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com