



VIEW LAB MONTHLY SUMMARY

Lab Name	Dowl.PHC, Pappalpetti	Details About Robust Analysis
Lab No	12600	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2023	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/09/2023

Date of Report Published : 00/10/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	SI
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosik (Semi automatic / Fully Automatic)	1870	810.86	4.92	14.46	808 mg/dL	-0.14	0.69
2	CREATININE II	Jaffes kinetic - Alkaline Picrate II	Robosik (Semi automatic / Fully Automatic)	1832	1.29	8.14	9.10	1.3 mg/dL	9.10	0.00
3	CHOLESTROL II	CHOD-PAP II	Robosik (Semi automatic / Fully Automatic)	1783	113.92	3.00	4.32	173 mg/dL	0.32	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out