

CAL	IBRATIO	N CERT	TIFICAT	T <u>E</u>	
CERTIFICATE NO: SBS/CL/23/12983		MEDICAL DEVICES		Page No:1 of 1	
Issue Date	06-10	0-2023			
SRF No & Date	SRF	23/00617-000	05 & 05-10-20)23	
Receipt Date	05-10	0-2023			
Calibration Date	05-10	0-2023			
Calibration Due	04-10	0-2024			
Customer Name & Address					
GOVERNMENT PRIMARY HEALTH CENTR	E(MAILAM BLOC	<),			
RETTANAI-604306, VILLUPURAM DISTRICT					
	Details of Device	Under Calibra	tion (DUC)		
Description : SEMI AUTO ANALYZ	ER Make	Make & Model : ROBONIK & PRIETEST TO		T TOUCH	
Range : MULTI	Sr. N	0	:	AT4340211RBK	
Resolution : MULTI	Identi	fication No	:	NA	
DUC Condition : SATISFACTORY	Locat	ion		LABORATORY	
Enviro	nmental Conditions	& Calibration	Procedure D	etails	
Environmental Details Temperature:25.6 ° C		Relative Humidity 52% RH		52% RH	
Calibration Procedure No SBS/CP/MD/2	20	Calibration done at		ONSITE	
-	Reference	Standards Det	ails	•	
S.No Description	Make/ SI No:	Certifi		ertificate No	
1 Electrical Safety Analyser Rigel Medical & 4		-1059	M-230809-16-4		10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±) Uncertainty in % (±)	
1	Insulation Resisitance	Measured values in MΩ		
	>20MΩ	97	13.92	
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)	
	<5000µAfor B,BF,CF	125	8.0	
	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)	
	<500µAfor B,BF,CF	235	7.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- $2. The \ user \ should \ determine \ the \ suitability \ of \ the \ instrument \ for \ its \ intended \ use.$
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

 Calibrated By

(Calibration Engineer)

M.DINESH





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS