

THE PIPETEMANN

NABL ACCREDITED CALIBRATION LABORATORY
ADDING ACCURACY TO YOUR EQUIPMENT

MASS - VOLUME - THERMAL - SPEED - PRESSURE - TIMER - ELECTRO TECHNICAL



CALIBRATION CERTIFICATE

CUSTOMER ADDRESS : M/s. THE MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
RAMANATHAPURAM
COIMBATORE
Phone No.:

UNIQUE LAB REPORT NUMBER : ULR-CC231723000000304F	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT : MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : TPMCC230156015	DATE OF ISSUE : 23-01-2023
RECEIPT NO : 11516	TEMPERATURE : 23.6 °C
RECEIPT DATE : 19-01-2023	HUMIDITY : 52 %
JOB CARD NO : TPM410012354438	AIR PRESSURE : 962 hpa
JOB CARD DATE : 19-01-2023	MEDIUM OF WEIGHING : GRADE III WATER
BRAND : FINNPIPETTE - F3	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 10 - 100 µl	DENSITY CORRECTION FACTOR [Z] : 1.0033 µl/mg
SERIAL NO : RW01615	CALIBRATION METHOD : ISO 8655:2002 PART - 6
TIPS USED : 200 µl	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE (1 µl - 10000 µl)	

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT : SARTORIUS
SL.NO : 19402655
RESOLUTION : 0.00001 g upto 80 g & above is 0.0001 g
CALIBRATED BY : THE PIPETEMANN(MASS),COIMBATORE
CERTIFICATE NUMBER : TPMCCM22050373
CALIBRATION VALIDITY : 03-05-2023
TRACEABILITY : NATIONAL STANDARDS