



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Subramanyar Street, Saidapet  
 Lab No: 9295  
 Month: May  
 Year: 2023  
 Constituent Group: Chemistry III

Details About Robot Analyser  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered: 19/05/2023

Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SI	SI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	202 mg/dL	2.25	0.44
2	CREATININE III	Jaffes Kinetic Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	111 mg/dL	0.62	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Study Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by  
**Swiftware Solutions**

Best Viewed at pixels 1024

*f. shall*  
 MEDICAL OFFICER  
 UPHC SAIDAPET,  
 VELLORE CITY MUNICIPAL  
 CORPORATION

41  
321



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Subramaniam Street, Saidapet  
 Lab No: 9295  
 Month: July  
 Year: 2023  
 Constituent Group: Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Yrse Value	SDI	SI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	103 mg/dL	-1.05	0.28
2	CREATININE III	Jaffes Kinatic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.3 mg/dL	1.34	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	110 mg/dL	-0.64	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by  
**Swiftware Solutions**

Best Viewed at pixels 1024

*J. Shall*  
 MEDICAL OFFICER  
 UPHC SAIDAPET,  
 VELLORE CITY MUNICIPAL CORPORATION



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Subramanyar Street, Saidapet  
 Lab No: 9295  
 Month: April  
 Year: 2023  
 Constituent Group: Chemistry II

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

S.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	SI
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robonk (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	137 mg/dL	2.43	0.34
2	CREATININE II	Jaffes kinetic - Alkaline Phorate II	Robonk (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL II	CHOD-PAP II	Robonk (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	101 mg/dL	1.83	0.28

GDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by  
 Swiftware Solutions

Best Viewed at pixels 1024 x 768

*[Signature]*  
 MEDICAL OFFICER  
 UPHC SAIDAPET,  
 VELLORE CITY MUNICIPAL  
 CORPORATION



MONTHLY SUMMARY

Site: Govt. UPHC, Subramaniam Street, Saidapet  
 ID: 9295  
 Month: January  
 Year: 2023  
 Analyte Group: Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 20/01/2023

Date of Report Published : 09/02/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	L
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1890	119.87	6.38	7.65	103 mg/dL	-2.20	0.3
UREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1865	2.21	10.02	0.22	1.9 mg/dL	-1.40	0.0
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1794	112.14	6.03	6.76	124 mg/dL	1.75	0.1

SDI Range	Interpretation
1 -1.00 to +1.00	Excellent.
1 ±1.01 to ±2.00	Good.
1 ±2.01 to ±2.99	Accept with caution. Warning Signal.
1 ±3.0	Unacceptable performance. Action Signal.

<a href="#">Home</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Site Maintained by  
**Swiftware Solutions**

Best Viewed at pixels 1024

*D. Small*

**MEDICAL OFFICER**  
**UPHC SAIDAPET**  
**VELLORE CITY MUNICIPAL CORPORATION**

2023 15:11:26

50377

96358



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Subramanyar Street, Saidapet  
 Lab No: 9295  
 Month: March  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robot Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered: 20/03/2023

Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	102 mg/dL	1.54	0.31
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	112 mg/dL	0.18	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Site Maintained by **Swiftware Solutions**

Best Viewed at pixels 1024

*d. shall*

MEDICAL OFFICER  
 UPHC SAIDAPET  
 VELLORE CITY MUNICIPAL CORPORATION

2.925044  
79.1396334

**CHRISTIAN MEDICAL COLLEGE**  
Vellore  
*The Mother of the Indian States 1900*

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Subramanyar Street, Saidapet  
Lab No: 9295  
Month: February  
Year: 2023  
Constituent Group: Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered: 21/02/2023

Date of Report Published: 08/03/2023

S.No	Analyte	Method / Reagent Name	Analyser Name	No of Participants	CV	Participants		Unit	SDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	158 mg/dL	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.1 mg/dL	0.36
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	110 mg/dL	-0.47

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Weekly Summary</a>	<a href="#">Change Password</a>	<a href="#">Log Home</a>	<a href="#">Sign-out</a>

Site Maintained by  
**Swiftware Solutions**

Best Viewed at pixels

*D. Chell*  
MEDICAL OFFICER  
UPHC SAIDAPET,  
VELLORE CITY MUNICIPAL  
CORPORATION