

| | | CALIE | BRA' | TION | CERT | IFICAT | <u>E</u> | |
|-------------------------------------|--------------------|-------------------|--------------------------------|------------------------|----------------|-------------------------|-------------|----------------|
| CERTIFICATE NO: SBS/CL/23/134 | | | /13406 | | MECHANICAL | | | Page No:1 of 1 |
| Issue Date | | | 07-10-2023 | | | | | |
| SRF No & Date | | | SRF/23/00741-0002 & 07-10-2023 | | | | | |
| Receipt Date | | | 07-10-2023 | | | | | |
| Calibration Date | | | | 07-10-2023 | | | | |
| Calibration Due | | | | 06-10-2024 | | | | |
| Customer Name & Ac | Idress | | | | | | | |
| GOVERNMENT PRIM | ARY HEALTH | CENTRE, | | | | | | |
| OLAPALAYAM-63805 | 2,KARUR DIST | RICT. | | | | | | |
| | | Deta | ails of D | evice Un | der Calibrat | ion (DUC) | | |
| Description : CENTRIFUGE | | | Make & | Model | : | KEMI & NA | | |
| Range : 3500 RPM | | Sr. No | | : | 29177 | | | |
| Resolution : 1 RPM | | Identification No | | : | NA | | | |
| DUC Condition : Satisfactory | | Location | | : | LABORATORY | | | |
| | Env | ironmental C | ondition | ns & Star | ndard Opera | ting Procedu | ire Details | |
| Environmental Details Temperature | | | 25.49 | : 25.4°C Relat | | midity | 54% Rh | |
| Calibration Procedure No SBS/CP/ML/ | | 04 Calibratio | | Calibration of | done at ONSITE | | | |
| | | | Refer | rence Sta | andards Deta | ails | | |
| S.No Description | iption | | | Make/ SI No: | | Certificate No | | Validity |
| 1 Digital Tachome | Digital Tachometer | | | LINE SEIKI / 175-0034V | | JRPM-CCTR-A&S-2023-0013 | | 09-06-2024 |

CALIBRATION RESULTS

| S.No DEVICE UNDER CALIBRATION | | STANDARD INSTRUMENTS | DEVIATION | EXPANDED UNCERTAINTY | |
|-------------------------------|------|----------------------|-----------|----------------------|--|
| | RPM | RPM | RPM | % | |
| 1 | 1000 | 998.9 | 1.1 | 4.2 | |
| 2 | 2000 | 1998.6 | 1.4 | 4.2 | |
| 3 | 3000 | 2998.7 | 1.3 | 4.2 | |

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By
(Calibration Engineer)

K.SATHYAMOORTHY



| Authorise | ed Signatory |
|-------------------------------|----------------|
| Quality Manager (C.SIVABALAN) | Chief Executiv |

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.