

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Boothalaipuram
 Lab No: 12996
 Month: September
 Year: 2023
 Constituent: Chemistry III
 Group:

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result
 Date of Result Entered: 19/09/2023
 Date of Report Published: 05/10/2023

SL No	Analyte	Method / Principle	Analyzer Name	No. of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	326 mg/dL	1.01	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	119 mg/dL	1.25	0.20
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

- Modify Address
- Modify Methods
- Subnet Lab Results
- View Lab Monthly Summary
- Yearly Summary
- Change Password
- Home
- Sign-out

67

September - 2023

Batch No	Ex. Date	Value	Report Date	SDI	Interpretation
09122054	Apr. 24	326 mg/dL	19/9/23	1.01	Good
740231120	Oct. 2024	119 mg/dL	19/9/23	1.25	Good
B102208	3/2025	1.2 mg/dL	19/9/23	-0.86	Excellent