

CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 06.10.2023
		Calibration Valid: 05.04.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, PAKKANADU, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MICROLUX
II	Model	5 μ l - 50 μ l
III	Serial No	PHC/PN/MP/002
IV	Location	LABORATORY

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05 μ l	20 μ l	50 μ l
1	05.05	20.08	50.97
2	05.06	20.07	20.02
3	05.01	20.05	20.96
4	05.97	20.99	20.02
5	05.01	20.08	20.09
6	05.02	20.95	20.04

For Heart Beat Medical Systems,


(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 06.10.2023
		Calibration valid : 05.04.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, PAKKANADU, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMOSCIENTIFIC
ii	Model	FINNPIPETTE F3
iii	Value	10 µl -100 µl
iv	Serial No	QW13604

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date: 06.10.2023
		Calibration valid: 05.04.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, PAKKANADU, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMOSCIENTIFIC
ii	Model	FINNPIPETTE F3
iii	Value	100µl-1000µl
iv	Serial No	QW14208


Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,


(Authorized signatory)




MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 06.10.2023
		Calibration Valid: 05.04.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, PAKKANADU, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MICROLUX
II	Model	-----
III	Value	1000µl
IV	Serial No	PHC/PN/MP/001

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	1000µl
1	1000.05
2	1000.06
3	1000.01
4	999.97
5	1000.01
6	1000.02

For Heart Beat Medical Systems,


(Authorized signatory)

